



# UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE  
United States Patent and Trademark Office  
Address: COMMISSIONER FOR PATENTS  
P.O. Box 1450  
Alexandria, Virginia 22313-1450  
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/816,636	04/01/2004	Patrick T. Petrino	10040213-1	1047
57299	7590	12/04/2006		
AVAGO TECHNOLOGIES, LTD. P.O. BOX 1920 DENVER, CO 80201-1920			EXAMINER YU, MELANIE J	
			ART UNIT	PAPER NUMBER

1641

DATE MAILED: 12/04/2006

Please find below and/or attached an Office communication concerning this application or proceeding.

74

<b>Office Action Summary</b>	<b>Application No.</b> 10/816,636	<b>Applicant(s)</b> PETRUNO ET AL.	
	<b>Examiner</b> Melanie Yu	<b>Art Unit</b> 1641	

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

**Period for Reply**

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

**Status**

- 1) ☒ Responsive to communication(s) filed on 11 August 2006.
- 2a) ☒ This action is **FINAL**.                      2b) ☐ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

**Disposition of Claims**

- 4) ☒ Claim(s) 11, 12, 21-23, 26 and 37-40 is/are pending in the application.
- 4a) Of the above claim(s) \_\_\_\_\_ is/are withdrawn from consideration.
- 5) ☐ Claim(s) \_\_\_\_\_ is/are allowed.
- 6) ☒ Claim(s) 11, 12, 21-23, 26 and 37-40 is/are rejected.
- 7) ☐ Claim(s) \_\_\_\_\_ is/are objected to.
- 8) ☐ Claim(s) \_\_\_\_\_ are subject to restriction and/or election requirement.

**Application Papers**

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☒ The drawing(s) filed on 01 April 2004 is/are: a) ☒ accepted or b) ☐ objected to by the Examiner.  
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).  
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

**Priority under 35 U.S.C. § 119**

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All    b) ☐ Some \*    c) ☐ None of:
1. ☐ Certified copies of the priority documents have been received.
  2. ☐ Certified copies of the priority documents have been received in Application No. \_\_\_\_\_.
  3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

\* See the attached detailed Office action for a list of the certified copies not received.

**Attachment(s)**

- |   |   |
|---|---|
| 1) <input type="checkbox"/> Notice of References Cited (PTO-892)  | 4) <input type="checkbox"/> Interview Summary (PTO-413)<br>Paper No(s)/Mail Date: _____ |
| 2) <input type="checkbox"/> Notice of Draftsperson's Patent Drawing Review (PTO-948)                        | 5) <input type="checkbox"/> Notice of Informal Patent Application                       |
| 3) <input type="checkbox"/> Information Disclosure Statement(s) (PTO/SB/08)<br>Paper No(s)/Mail Date: _____ | 6) <input type="checkbox"/> Other: _____  |

### **DETAILED ACTION**

1. Applicant's amendment filed 11 August 2006 has been entered. Claims 11, 12, 21-23, 26 and 37-40 are currently pending in this application.

### ***Withdrawn Rejections***

2. Previous rejections under 35 USC 103(a) have been withdrawn in light of applicant's amendments and arguments.

### ***Claim Rejections - 35 USC § 102***

The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless –

(b) the invention was patented or described in a printed publication in this or a foreign country or in public use or on sale in this country, more than one year prior to the date of application for patent in the United States.

3. Claims 11, 23, 26 and 39 are rejected under 35 U.S.C. 102(b) as being anticipated by Daniels et al. (US 2002/0004246).

Daniels et al. teach a test system comprising: a medium containing a labeling substance that comprises first persistent fluorescent structures that emits light having a first frequency and second persistent fluorescent structures that emit light having a second frequency, wherein each of the first persistent fluorescent structures is attached to a substance that is capable of binding the first structure to a target analyte when a sample containing the target analyte is applied to the medium (each detection reagent is associated with a nanocrystal having a distinct emission peak and nanocrystal is a persistent fluorescent structure, par. 207; capture and control reagents bind the first and second structures and are present in a chromatographic medium therefore the first and second nanocrystal structures are present in a medium, par. 201; emission peak of nanocrystal incorporated into control is distinct from that exhibited by nanocrystals of the first detection reagent therefore first and second nanocrystals emit at different frequencies, par. 198); a

Art Unit: 1641

light source positioned to illuminate a target area and a control area on the medium (par. 213); a first photodetector positioned to measure light of the first frequency originating from the target area of the medium (multiple detectors for each light emission frequency, par. 214); and a second photodetector positioned to measure light of the second frequency originating from the control area, wherein a signal from the second photodetector indicating an intensity above a threshold level indicates that the sample has passed through the target area (separate detector for each detection region with a different emission frequency, par. 214; control region has a different emission frequency than detection region, par. 198; detection of nanocrystals in the control region occurs in the presence or absence of analyte and therefore indicates that the sample has passed through the medium, par. 242).

With respect to claim 23, Daniels et al. teach the first and second persistent fluorescent structures comprising quantum dots (par. 198 and 79).

Regarding claim 26, Daniels et al. teach the medium comprising a lateral-flow strip for performing a binding assay (par. 200-201) and the target area containing an immobilized substance that binds to and holds the complex including one of the first persistent structures and the target analyte (par. 200-201; capture reagent binds to the detection complex, par. 189; detection complex comprises analyte and nanocrystal, par. 137-139; capture reagent is in a capture region, par. 115).

With respect to claims 39 and 40, Daniels et al. teach the second persistent structures bind to the control strip (control ligands are in a control region, par. 115; control ligands bind to second persistent structures that have an emission frequency different from that in the capture region, par. 198). Daniels et al. also teach a first and second color filter corresponding to the first and second photodetector that transmit the first and second frequencies, respectively (multiple detectors are present, one for each region, and each has

Art Unit: 1641

a bandpass filter for detecting a narrow wavelength range corresponding to the nanocrystal emission wavelength in the capture and control regions, par. 214).

***Claim Rejections - 35 USC § 103***

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

The factual inquiries set forth in *Graham v. John Deere Co.*, 383 U.S. 1, 148 USPQ 459 (1966), that are applied for establishing a background for determining obviousness under 35 U.S.C. 103(a) are summarized as follows:

1. Determining the scope and contents of the prior art.
  2. Ascertaining the differences between the prior art and the claims at issue.
  3. Resolving the level of ordinary skill in the pertinent art.
  4. Considering objective evidence present in the application indicating obviousness or nonobviousness.
4. Claims 12, 21, 22, 37 and 38 are rejected under 35 U.S.C. 103(a) as being unpatentable over Daniels et al. (US 2002/0004246) in view of Crosby (US 6,217,744).

Regarding claims 37 and 38, Daniels et al. teach a rapid diagnostic test system comprising a medium, first photodetector and second photodetector, but fail to teach the first and second photodetectors and medium contained in a single-use module that can be inserted into a reusable module.

Crosby teaches a photodetector and medium necessary for optical detection contained in a single use module (optical components and porous membrane are part of the disposable device, optical components comprise the photodetector, col. 5, lines 64-col. 6, line 13) and inserted into a reusable module for communication of test signals between the single-use module and reusable module (communication between disposable analysis device

Art Unit: 1641

and information gather and storage system, disposable device is single use and information gathering system is reusable, col. 6, lines 57-67; col. 7, lines 37-45), in order to provide a self powered device that resists corrosion and degradation.

Therefore it would have been obvious to one having ordinary skill in the art at the time the invention was made to include in the test system of Daniels et al., the photodetectors and medium necessary for optical detection contained in a single use module that can be inserted into a reusable module for communication of test signals as taught by Crosby, in order to provide small, point of care diagnostic tests that are small in size and produces a fast quantitative or qualitative result with increased reliability. Although Crosby does not specifically teach two photodetectors, it would have been obvious to include all photodetectors of Daniels et al. necessary for detection in the single use device taught by Crosby.

With respect to claims 12 and 21, Crosby teaches the reusable module implementing a user interface capable of indicating a test result on a display (console is the information gathering and storage system and has a display screen to display results from the disposable device, col. 7, lines 37-50).

Regarding claim 22, Crosby teaches that the test signals are electrical signals (col. 7, lines 14-25).

### ***Response to Arguments***

5. Applicant's arguments with respect to claims 11, 12, 21-23 and 26 have been considered but are moot in view of the new ground(s) of rejection. Previous rejections under 35 USC 103(a) have been withdrawn. Upon further consideration, a new ground(s) of rejection is made in view of applicant's amendment requiring a first and second photodetector and first and second persistent fluorescent structures.

***Conclusion***

No claims are allowed.

6. Applicant's amendment necessitated the new ground(s) of rejection presented in this Office action. Accordingly, **THIS ACTION IS MADE FINAL**. See MPEP § 706.07(a).

Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

A shortened statutory period for reply to this final action is set to expire THREE MONTHS from the mailing date of this action. In the event a first reply is filed within TWO MONTHS of the mailing date of this final action and the advisory action is not mailed until after the end of the THREE-MONTH shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of the advisory action. In no event, however, will the statutory period for reply expire later than SIX MONTHS from the date of this final action.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Melanie Yu whose telephone number is (571) 272-2933. The examiner can normally be reached on M-F 8:30-5.


If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Long Le can be reached on (571) 272-0823. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Art Unit: 1641

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.



Melanie Yu  
Patent Examiner  
Art Unit 1641



LONG V. LE

11/13/06

SUPERVISORY PATENT EXAMINER  
TECHNOLOGY CENTER 1600